

APR 28 2010

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130179</u>		
Supervisor/Operator Performing the Verification Procedure:		
Name <u>V. Eugene Fowler</u>	ID# <u>4386</u>	Date <u>4-26-10</u>
A Agency <u>Alaska State Troopers</u>		Phone # <u>262-4453</u>
Instrument Location <u>Soldotna AST Post</u>		
B Alco S/N <u>X124630</u> Target Value <u>.079</u> High Pressure <u>800</u>		
Alco Test Value Average	<u>.076</u> 1 st Alco	<u>.076</u> 2 nd Alco
Signature <u><i>V. E. Fowler</i></u>		
(OVER)		

BMB
5/3/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 8th day of June, 2010.

Carolyn M. Noland
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

APRIL 26, 2010

OPERATOR'S NAME:

FOAMER/ONE

OPERATOR'S NUMBER: 4386

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.I.L. # : UOC

DEPT/AGENCY: SLTS

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .079

ALCO S/N: X124630

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.47 in
ALCO TARGET .077 16:20
BLANK TEST .000 16:21
INTERNAL STANDARD VERIFIED 16:21
ALCO TO 29.47 in .076 16:22
BLANK TEST .000 16:23
SUBJECT SAMPLE .000 16:23
BLANK TEST .000 16:24
ALCO TO 29.47 in .076 16:24
BLANK TEST .000 16:25

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

APRIL 26, 2010

TIME 15:27

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/28/01

HEATERS

SAMPLE CHANGER: 43c

BREATH TUBE: 42c

BAROMETER: 29.47 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

11#1V/CX++-1-0107436789:1K=V9980DEF6
HIJKLWOPQRSTUWXYZ1234567890
pqrstuvwxy1234567890